

# (12) UK Patent Application (19) GB (11) 2 391 625 (13) A

(43) Date of A Publication 11.02.2004

(21) Application No: 0218547.8

(22) Date of Filing: 09.08.2002

(71) Applicant(s):  
**Diagnostic Ultrasound Europe B.V.**  
(Incorporated in the Netherlands)  
Lage Dijk 14, 3401 RG IJsselstein,  
Netherlands

(72) Inventor(s):  
**Henri L G Baartmans**  
**Gerald James McMorow**  
**Nicolaas Bom**  
**Charles Theodoor Lancee**

(74) Agent and/or Address for Service:  
**Eric Potter Clarkson**  
Park View House, 58 The Ropewalk,  
NOTTINGHAM, NG1 5DD, United Kingdom

(51) INT CL<sup>7</sup>:  
**A61B 8/00**

(52) UK CL (Edition W ):  
**G1G GMD**

(56) Documents Cited:  
**EP 0271214 A2** **WO 2000/071030 A1**  
**US 6359190 B1** **US 6213949 B1**  
**US 5964710 A** **US 4926871 A**

(58) Field of Search:  
UK CL (Edition T ) **G1G**  
INT CL<sup>7</sup> **A61B**  
Other: **Online: WPI, EPODOC, JAPIO**

(54) Abstract Title: **Instantaneous ultrasonic echo measurement of bladder urine volume with a limited number of ultrasound beams**

(57) In a first version the apparatus is composed of a transducers assembly that transmits a plurality of narrow ultrasound beams in different directions towards the bladder and receives the returning ultrasound signals; a receiver detector for processing the returned signals; an analog-to-digital converter; a memory to store the digitized data and a volume display allowing to define the optimal position of the transducer assembly. The apparatus also includes a signal processing software that automatically determines the bladder Depth D and Height H and computes the volume of urine using an empirical formula corrected by specific, empirically measured, filling dependant correction factors. In a second version a single wide angle ultrasound beam transducer transmitting ultrasound signals at fundamental frequency is used to quantify the urine volume. Return signals originating from a depth beyond the usual position of the posterior wall depth of a filled bladder are analyzed for presence of higher harmonic signals which in turn are related to presence or absence of urine. Both methods or a combination thereof can be used as a simple warning device for presence of residual urine after voiding or indicate the presence of a critical bladder urine filling level.




Figure 1

GB 2 391 625 A




Figure 1





Figure 2. Various, filling dependent measurement configurations.







Figure 4

Figure 5






Figure 6







Figure 7



Figure 8










Figure 11

**INSTANTANEOUS ULTRASONIC ECHO MEASUREMENT OF**  
**BLADDER VOLUME WITH A LIMITED NUMBER OF**  
**ULTRASOUND BEAMS.**

**5 TECHNICAL FIELD**

This invention concerns an apparatus which, in a **first version**, with a limited number of fixed ultrasound transducers with narrow sound beams oriented in well defined directions, automatically determines the volume of the human bladder without assumption of any geometrical bladder shape,  
10 where volume is calculated by (Height x Depth x K) and the empirically measured K factor varies with bladder filling degree, which in turn is indicated by the number of ultrasonic beams that intercept the filled bladder. In this first version, standard echographic technique is used where short ultrasound pulses are transmitted at fundamental frequency and the echo  
15 travel time is used to calculate distance.

In a **second version**, with a wide ultrasound beam, pulses are transmitted at fundamental frequency. Due to the wide sound beam this beam encompasses a large part the volume of a possibly filled bladder. Echo  
20 signals from a large distance W, where W is the average distance from the transducer in dorsal direction to a point beyond the posterior wall of an average filled bladder, are analyzed for higher harmonic contents. Non-linear behavior will increase with depth and particularly be stimulated by presence of urine. Attenuation of returned echo signals from a large distance  
25 will be considerably less in the presence of urine. A combination of these two effects will favor presence of higher harmonics as compared to the presence of the fundamental frequency in the return signal. With this

information urine quantity or a critical urine filling level of the bladder can be established.

In a **third version** a combination of a narrow ultrasound beam for detection  
5 of the posterior bladder wall distance  $W$  with the wide acoustic beam approach for subsequent measurement of urine filling of the bladder is described.

### **BACKGROUND OF THE INVENTION**

10 It is well known that bladder dysfunction is associated with a number of clinical conditions requiring treatment. In many of these cases it is important to accurately determine the volume of the bladder. Under other conditions such as post-operative recovery, where there is temporary loss of bladder sensation and /or loss of the normal voiding mechanism too much  
15 distention of the bladder has to be avoided. Under those conditions voiding by catheter introduction is carried out. However, serious disadvantages to unnecessary catheterization range from the uncomfortable situation for the patient to serious possibilities of infection. Thus, a non-invasive quick measurement of bladder volume with the patient usually in the supine  
20 position is indicated. Sometimes the accurately determination of volume is indicated, sometimes however an indication is sufficient. Questions that may be asked are for instance after voiding: "is there still too much urine left?" or after surgery "is the bladder filling above a certain level so that voiding is necessary?"

25

Non-invasive procedures for bladder volume estimation are known, but are either unreliable or expensive or have some other significant disadvantages. Palpation and auscultatory percussion are known to be unreliable, while radiography and dye-excretion techniques are known to be similarly

(  
inaccurate. For assessing bladder volume, catheterization remains the “gold standard” however it is invasive, painful and might produce traumas or infections.

## 5    **SUBJECT**

The described technique concerns measurement of urine volume in the humane bladder with the use of pulsed ultrasound with a limited number of ultrasound transducers.

10    **In a first version** a limited number of transducers are mounted in a transducer assembly. The assembly is positioned non-invasively at the body skin over the position of the bladder with the patient in a supine position. For acoustic contact a coupling gel may be used. Each ultrasound transducer in the assembly transmits and receives the ultrasound signal in a narrow  
15    beam through the contact plane. During the measurement the transducers are used in a certain succession. All transducers have been mounted in the assembly such that in transmission and reception successively the beams penetrate the area of the bladder in approximately the sagittal cross sectional plane. The sagittal plane is here defined as antero-posterior plane of the  
20    body. One transducer beam direction is dorsal with in addition at least one transducer beam in the dorsal-caudal and one transducer beam in the dorsal-cranial direction. The volume is calculated on the basis of two bladder measurements defined in the sagittal plane as Depth (D) and Height (H). These measurements are derived on the basis of echo travel time from  
25    echoes originating at the anterior and posterior bladder wall. Depth is in principle a measurement in dorsal direction. Height is a measurement approximately in the cranial direction. The volume is calculated depending on the specific, filling dependent, measurement configuration following the formula  $D \times H \times K$ . Where K is an empirically measured, filling configuration

(  
dependant, correction factor. Beam directions and examples for D and H are illustrated in figures 1 and 2.

In a second version of the described technique a single wide beam  
5 ultrasound transducer is positioned non-invasively at the body skin over the  
location of the bladder. The wide beam can be created by the curved surface  
of the transducer or by a flat acoustically active surface of for instance a disk  
shaped transducer supplied with a curved lens. Ultrasonic signals are  
transmitted and received in the wide, cone like, ultrasound beam and  
10 propagation is approximately spherical. Similar to the above described  
method a pulsed echo signal is transmitted at fundamental ultrasonic  
frequency. In this second version of the described technique echo data are  
analyzed as originating from a distance beyond the average position of the  
posterior (filled) bladder wall. The received echo signal will contain  
15 information over almost the entire bladder as encompassed by the wide  
ultrasound beam. Due to non-linearity, higher harmonic components will  
build up during propagation and thus be reflected in the returning echo.  
Compared to propagation through normal tissue, the presence of higher  
harmonics in the signal is greatly stimulated when propagating through  
20 urine. Analyses of presence of higher harmonic components in relation to  
the fundamental frequency is used for indication of presence of urine in the  
bladder. Neutralizing patient variation as to obesity etc can also be  
accomplished by comparing echo signals received from sequentially  
transmitted pulses at low transmit power (linear propagation only) and pulse  
25 transmission at high power (enhancing non-linearity).

## **PRESENT LEVEL OF THE TECHNIQUE**

Non-invasive bladder volume measurement techniques with ultrasound  
echography have been described. In principle, echography measures



(

distance based on echo-travel-time. Early echo techniques did use a single ultrasound transducer and echo presentation was recorded as echo amplitude versus depth. West, K A: "Sonocystography: A method for measuring residual urine", Scand J Urol Nephrol 1: pp68 – 70, 1967 describes the subsequent use of some discrete beam directions. He does not have a separate transducer for each beam direction. His method is only qualitative, not instantaneous, and based on distance measurement to the dorsal posterior bladder wall. His method is not adjusted to specific, filling dependant, measuring configurations. A relation between the difference in echo travel time between echoes from the posterior an anterior bladder wall and the independently measured bladder volume has been reported by Holmes, J H: "Ultrasonic studies of the bladder", J Urology, Vol 97, pp.654 – 663. His described volume measurement method is exclusively based on bladder depth measurement. Since the bladder changes in shape when filling, a single distance measurement is not precise enough to predict the entire bladder volume. No filling dependent measurement configuration is used.

Diagnostic ultrasound is today well known for real-time cross-sectional imaging of human organs. For cross-sectional imaging the sound beam has to be swept electronically or mechanically through the cross section to be imaged. Echoes are presented as intensity modulated dot on the display. The instruments are costly and require a skilled operator. Volume is sometimes calculated based on bladder contours obtained in two orthogonal planes with a geometric assumption of bladder shape. For 3-dimensional or volumetric echography the sound beam has to be swept through the entire organ. This further increases complexity, acquisition time of the data, and costs of the instrument.

Hakenberg *et al.*: "The Estimation of Bladder Volume by Sonocystography", J Urol, Vol 130, pp249 – 251, have reported a simple method that is based on measuring the diameters obtained in a cross sectional image in the midline sagittal bladder plane only. The bladder volume has been related to  
5 bladder Height and Depth as follows: Volume is Height x Depth x 6.6 ml. This formula showed a good correlation coefficient ( $r=0.942$ ) with a relatively large average error of 30.1%. For this approach a two-dimensional imaging apparatus was required. The used apparatus is complex and is different from the method described in this application. It does not use a  
10 single wide beam transducer or a limited number of fixed transducers in an assembly or a combination of this.

An ultrasound apparatus for determining the bladder volume is shown in U.S. Pat. No. 4,926,871 in the name of Dipankar Ganguly *et al.* In this text,  
15 a number of possibilities are mentioned, amongst which a scan head embodiment referred to as a sparse linear array with transducers mounted at predetermined angles with sound beams pointing towards the same position. The volume is calculated according to a geometric model. In the claims an apparatus is described, involving an automatic calculation of bladder volume  
20 from ultrasound measurements in a first and second plane, which are substantially orthogonal to each other. Sound beams are deflected by a stepper motor. It requires a skilled operator to manipulate the scan head in a particular way to obtain the ultrasound measurements. For the volume calculation method described in this application no use is made of any  
25 geometrical model of the bladder, whereas only a limited number of sound beams approximately in the sagittal plane, or a single wide beam is used.

Volume measurement based on echographic sampling of the bladder with a hand guided transducer mounted in a panthograph has been described by

Kruczkowski *et al*: "A non-invasive ultrasonic system to determine residual bladder volume", IEEE Eng in Medicine & Biology Soc 10th Ann Conf, pp1623 – 1624. The sampling covers the entire bladder, follows a given pattern and is not limited to a single or two cross sections of the bladder. For the calculation he needs data from many beam directions. The acquisition procedure is time consuming and thus no instantaneous volume measurement results. The method described in this application is based on use of a single, wide beam or the use of a limited number of mutually fixed sound beams directions with instantaneous volume indication.

10

The hand steered transducer guiding for recording of echo data from the bladder has subsequently gained in acquisition speed by introduction of constructions whereby the transducer, and thus the beam, was mechanically swept. This nevertheless still requires an acquisition time equivalent to full acquisition procedure and thus does not yield an instantaneous display of volume. No instantaneous feedback on optimal positioning is thus available. An example of such methods is the *Bladderscan*. In the *Bladderscan* Technology (registered trademark of Diagnostic Ultrasound Corporation) bladder volume is measured by interrogating a three-dimensional region containing the bladder and then performing image detection on the ultrasound signals returned from the region insonated. The three dimensional scan is achieved by performing twelve planar scans rotated by mechanically sweeping a transducer through a 97 degree arc in steps of 1.9 degrees. The three dimensional scanning requirement makes this instrument complex. It can not be compared with the simple approach described in this application.

25

Yet another ultrasound method "System for estimating bladder volume" is described by Ganguly *et al* in patent 5,964,710 dated October 12, 1999. This

(  
method is based on bladder wall contour detection with echographically  
obtained data in a plurality of planes which subdivide the bladder. In each  
single plane of the plurality of planes a number of N transducers are  
positioned on a line to produce N ultrasound beams to measure at N  
5 positions the distance from front to back wall in the selected plan. From this  
the surface is derived. This procedure is repeated in the other planes as well.  
The volume is calculated from the weighted sum of the plurality of planes.  
In Ganguly's method the entire border of the bladder is echographically  
sampled in 3 dimensions. His method differs strongly from the method  
10 described in this application whereby only a single wide beam is used or a  
limited number of mutually fixed sound directions are used in approximately  
a sagittal plane with a filling dependent measurement configuration.

An important parameter for assessing bladder volume if this volume has to  
15 be derived from a limited number of beams or planes is the knowledge of  
bladder shape and position which can drastically vary with age, gender,  
filling degree and disease. In the adult patient the empty bladder has the  
shape of a triangular prism and is located behind the pubis. When it is  
progressively filled, there is first a distention of the bladder depth followed  
20 by an expansion of the bladder height. The bladder shape is complex and  
can not be represented by a single geometrical formula such as ellipsoid,  
sphere etc. This explains the large error that several studies obtained when a  
single geometric model was used. However there exists a correlation  
between the bladder height and the bladder widening with progressive  
25 filling.

In the first approach of the present invention an instrument is described  
which allows assessment of bladder volume by using only a few ultrasound  
beams appropriately oriented in approximately the sagittal plane. The

narrow sound beams in principle diverge relative to each other. This allows covering a wide range of filling degrees of the bladder, from almost empty, when the bladder is located behind the pubis, to a full bladder that causes a substantial bladder height (See figure 1 and 2). From each beam can be established, by detection of the posterior bladder wall echo, if this beam does pass a filled bladder. From the knowledge of all beams that do pass the filled bladder the appropriate filling or measurement configuration follows. The acoustic beams are positioned in such a way that the Depth D and Height H of the bladder can be estimated for the specific measurement configuration. The volume of urine is then computed from an empirical formula  $D \times H \times K$  that does not depend on any geometric model. K is a known, empirically established correction factor which is specific for each measurement configuration and has been established by calibrated bladder measurements on a prior series of patients. The accuracy of the first approach is thus based on an a priori known correction factor which is related to a specific filling degree, which in turn depends on the number of beams that intercept the filled bladder.

A second version of the instrument is based on the measurement of the presence of higher harmonics in the echo signal. For this approach the echo signal from a depth greater than the distance from the transducer to the posterior bladder wall must be analyzed. For a filled bladder in adults this depth W would be approximately 12 cm.

It is known that when sound pulses are transmitted at a fundamental frequency, higher harmonics of this fundamental frequency may be present in the received echographic signal. Non-linear distortion increases with distance, insonifying ultrasound energy and frequency. Attenuation diminishes the ultrasound amplitude with increasing propagation distance

and reduces the higher harmonic energy. Since attenuation of the ultrasound signal in urine is low compared to tissue and non-linear distortion in urine is large compared to tissue it results that urine is very different from tissue in its ability to generate higher harmonics. We have measured the presence of  
5 higher harmonics in the echo signal from 12 cm depth when the bladder was filled. With an empty bladder the echoes obtained from the same depth did not contain higher harmonics.

The interest of higher harmonic signals in the ultrasound technique stems  
10 from echo contrast technology. Echo contrast material contains coated gas containing micro bubbles suspended in a fluid. These bubbles can create higher harmonic components in the echo signal due to non-linearity. This is used to indicate presence of contrast on the diagnostic image. A wide variety of pulse techniques is used to stimulate echographic visibility of  
15 contrast. These include multi pulse procedures, multi frequency procedures, power Doppler imaging, pulse coding, pulse inversion and other imaging methods. A survey is documented in "Ultrasound Contrast Agents" ISBN 1-85317-858-4 chapter 3 "Contrast-specific imaging methods" by de Jong *et al.* With a single transducer with wide sound beam, such as results with a  
20 curved acoustic element or a flat, disk shaped transducer plus curved lens, the propagating sound beam would encompass the entire bladder. The transducer must be designed to optimally transmit the fundamental ultrasound frequency and at the same time be capable to receive fundamental and higher harmonic echo signals. Broadband piezo-electric ceramic  
25 transducers have been described as well as combination transducers using ceramic in transmission and PVDF material in reception. In transmission a single or multi pulse procedure can be followed. If the returned echo signal with such a method would, in relation to the fundamental echo signal, be

analyzed for the presence of higher harmonics, the presence of a certain level of bladder filling or the volume of urine can be established.

## **DESCRIPTION OF THE DRAWINGS**

5    **FIG. 1** Illustrates a sagittal (anteroposterior) cross sectional plane of a patient in supine position where a transducer assembly 1 with transducers A, B, C, D and E, is positioned on the abdominal wall just above the Symphysis Pubis 2 and the ultrasound beams are indicated to cross the area of the partially filled bladder 3. From the transducer assembly, the sound beam A intercepts  
10    the bladder area in dorso-caudal direction, soundbeam B intercepts the bladder in dorsal direction and sound beams C, D, and E respectively in dorso-cranial direction. In figure 1 the patient's leg is indicated by 4.

**FIG. 2** Illustrates various bladder filling stages from an almost empty bladder  
15    to a strongly filled bladder and the corresponding measurement configurations. Depth D and Height H have been defined for each filling situation as indicated and are calculated from detected bladder wall echoes taking the specific measurement configuration into account. For each measurement configuration a specific Depth D and Height H is defined.

20

**FIG 3.** Illustrates, by way of example for a transducer assembly with 5 transducers (here only A and D, necessary for calculation of H are shown) , the calculation of Height H (5) in the measurement configuration when bladder posterior wall echoes are detected originating from sound beam A,  
25    B, C and D. This is the "filled bladder" measurement configuration shown in figure 2. Apparently no posterior wall echoes are detected in sound beam E because the bladder filling is not yet in a strongly filled stage and thus beam E does not intercept the bladder. Depth D is derived from beam B (not shown in figure 3).

FIG. 4 Represents a flow chart of the principle hardware components. In this block diagram a “useful” transducer signal occurs when bladder wall echoes are detectable in its sound beam.

5

FIG 5. Illustrates a top view of 5 disk shaped transducers in a possible transducer assembly. The distance between transducers B, D and C, A, E and their positioning is such that all sound beams can be assumed to be in approximately a sagittal cross section through the bladder. Yet another transducer assembly with 4 transducers in a row is also illustrated.

10

FIG 6. Illustrates a cross sectional view showing in the length direction a possible transducer and related sound beam orientation when 5 single transducers are used.

15

FIG. 7 Illustrates the sagittal cross sectional plane with a single wide beam transducer non-invasively positioned on the abdominal skin surface over the filled bladder 3. Echo signal is received from a range at depth W.

20 FIG 8. Illustrates the measured received scattered power in the fundamental frequency  $f_0$  and the higher harmonic frequencies  $2f_0$  and  $3f_0$  in a situation with an empty versus a filled bladder.

FIG 9. Is a block diagram illustrating the principle components of the bladder volume measurement instrument based on a single ultrasound wide beam where detection of presence of higher harmonics in the received signal from a give depth range is used to measure volume. Two different transmit levels are used to enhance the bladder effect and eliminate patient variation.

25



FIG. 10 shows two possible transmit pulse sequences to enhance the difference between linear and non-linear sound propagation.

FIG. 11 Illustrates a possible look-up table based on prior calibrated patient bladder volume measurements relating presence of harmonic power in the received echo signal versus volume.

#### **A) DETAILED DESCRIPTION OF THE FIRST METHOD.**

The first method describes a simple device that allows the assessment of bladder volume, using only a few beams appropriately oriented. Under the assumption that there exists a correlation between the bladder height and width, a simple approach has been developed. It consists of a limited number of acoustic beams positioned in such a way that the depth D and the height H of the bladder could be estimated in approximately a single sagittal plane. The volume of urine is then computed from an empirical formula that does not assume any geometric model.

In operation of the apparatus of the present invention, the transducer assembly 1 is placed on the abdomen of the patient in the supine position, just above the symphysis pubis 2. We are presenting a particular configuration of the assembly 1. Nevertheless, various configurations can be derived from this model and several modifications could be achieved (number of transducers, position, orientation, etc...) without departing from the initial ideas. The device proposed as an example is composed of 5 disc shaped transducers A, B, C, D and E (focused or non-focused) positioned in the assembly at predetermined distance from each other (Figure 5 top panel) and oriented at predetermined angles  $\theta_A$ ,  $\theta_B$ ,  $\theta_C$ ,  $\theta_D$  and  $\theta_E$  (Figure 6). Referring to Figure 5 (top panel), it appears that the transducers A, B, C, D

and E are oriented in two different planes. The distance between these two planes is small compared to the bladder 3 size and thus we can assume that the information received from each transducer represent the characteristics of approximately a single sagittal or anteroposterior plane. The orientation of each beam has been determined from the knowledge of the bladder 3 position and shape when it is filling up as measured in a patient series. The first beam of the transducer assembly 1 (soundbeam from transducer A) is oriented in such a way that it reaches the bottom of the bladder, passing just above the symphysis pubis 2. The remaining beams are positioned for successively intercepting the bladder 3 when it expands with increasing filling degree.

Computation of the Depth D and Height 5: Depending on the number of beams that are intercepting the bladder 3 and on the geometrical configuration of the transducer assembly (1), the distances H and D are determined by different mathematical procedures. For most measurement configurations the depth D of the bladder is determined by the distance between echoes derived from front and back wall of the bladder estimated from Transducer B.

The Height H (5) calculation in the specific measurement configuration (here we selected as an example the “ filled bladder” configuration of figure 2) when posterior bladder wall echoes are detected in signals obtained in beam A, B, C, and D, but not in beam E is illustrated in figure 3. For the other filling geometries the height is calculated in a corresponding way. The mathematical procedure is as follows:

$$\cos \theta_A = \frac{[AA2]}{[AA1]} \Rightarrow [AA2] = \cos \theta_A \cdot [AA1] \quad (1)$$

$$\sin \theta_A = \frac{[A1A2]}{[AA1]} \Rightarrow [A1A2] = \sin \theta_A \cdot [AA1] \quad (2)$$

$$\cos \theta_D = \frac{[DD2]}{[DD1]} \Rightarrow [DD2] = \cos \theta_D \cdot [DD1] \quad (3)$$

$$\sin \theta_D = \frac{[D1D2]}{[DD1]} \Rightarrow [D1D2] = \sin \theta_D \cdot [DD1] \quad (4)$$

$$5 \quad [A1I] = |[AA2] - [DD2]| \quad (5)$$

$$[ID1] = [D1D2] + [A1A2] + [AD] \quad (6)$$

$$\Rightarrow Height = [A1D1] = \sqrt{[A1I]^2 + [ID1]^2} \quad (7)$$

- 10 Volume computation: The volume of urine is correlated to the bladder diameter (Height 27 and Depth 26) by the empirical formulae:

$$Height * Depth * K$$

- 15 where K is a correction factor. Depending on the number of beams that allow the determination of the bladder dimensions (from 1 to 5) and others parameters such as the age, the gender, the correction factor is different. For a given situation (parameters other than number of beam are fixed), the correction factors K1, K2, K3, K4 and K5 are optimized using linear  
20 regression analysis.

The hardware configuration is illustrated in the flow chart of figure 4.

- 25 After positioning the transducer assembly correctly over the bladder area the measurement procedure is started by pressing the start button which during the (short) measurement procedure remains depressed. Subsequently the transducers are activated for transmission of ultrasound pulses and reception

of echoes and possible detection of bladder wall echoes in a specific order. Thereafter it is established, when a clear posterior bladder wall echo is detected, which ultrasound beams, this we call here the beams of “useful” transducers, penetrate the filled bladder. From this the filling situation or measurement geometry is established. As a result the proper correction factor can be selected. After calculation of the volume the value is stored in memory and displayed. During the measurement procedure the transducer assembly is slightly moved and memory data are refreshed if a larger volume is measured. The highest value will correspond with the correct bladder volume. This is displayed.

The described **first method** differs greatly from known other apparatus:

- 1) The device is composed of a limited number of static single element transducers;
- 2) The arrangement of the transducer is not similar to the arrangement of a linear array;
- 3) The transducers are oriented towards the bladder with specific angles allowing the estimation of the urine volume over a wide range of volumes;
- 4) The method for automatic volume computation does not assume any geometrical model for the bladder shape;
- 5) It is valid for any bladder shape since the volume is computed with an empirical formula for various filling ranges;
- 6) It is not based only on the measurement of distances between the front and back wall or area in different planes;
- 7) It uses an automatic detection of the bladder height and depth depending on the number of beams that intercept the bladder;

- (
- 8) It optimizes the correction factor depending on the degree of filling (or other factors, such as age, gender, weight, that may influence the calculations);
  - 9) The device includes a close loop to easily find the optimal position;
  - 10) The optimal position corresponds to the largest volume computed;
  - 11) The device works instantaneously.

## **B) DETAILED DESCRIPTION OF THE SECOND METHOD.**

10 The second version of the device is based on a different principle. The approach consists of using a single acoustic beam with a very wide width such that it encloses approximately the entire volume of the bladder when it is filled up. Such a wide beam width can be obtained using a single element transducer with a defocusing lens as drawn in figure 7 or a curved single  
15 element transducer. The schematic principle of transducer positioning is illustrated in figure 7. The sagittal cross section through the bladder is shown. The cone like shape of the acoustic beam allows to encompass approximately the full bladder volume, and therefore any harmonic distortion detected in the echo signal returning from a region beyond the  
20 posterior wall of the bladder around depth  $W$ , would correlate to the amount of fluid contained in the bladder. Indeed, it has been demonstrated that the propagation of ultrasound waves is a nonlinear process. The nonlinear effects, which increase with higher intensities, have been predicted and demonstrated at frequencies and intensities used in the diagnostic range  
25 either in water or in human body (A Baker *et al*: "Distortion and high-frequency generation due to non-linear propagation of short ultrasonic pulses from a plane circular piston", J Acoustic Soc Am 92(3), pp1699 – 1705). The distortion is due to slight non-linearities in sound propagation that gradually deform the shape of the propagating sound wave, and result in

development of harmonic frequencies which were not present in the transmitted wave close to the transducer. This manifests itself in the frequency domain as the appearance of additional harmonic signals at integer multiples of the original frequency. These effects occur most strongly when

5 ultrasound waves propagate within liquids with relatively low acoustic attenuation such as water, amniotic fluid or urine. Indeed, acoustic propagation in fluids gives rise to extreme nonlinear effects at diagnostic frequencies. Within soft tissues, nonlinear processes also take place but are modified as a result of the different acoustic characteristics of these tissues,

10 most notably their high acoustic absorption. Indeed, water and amniotic fluids (urine) are significantly different from tissue. It is known from literature (A C Baker: "Prediction of non-linear propagation in water due to diagnostic medical ultrasound equipment", *Phys Med Biol* 1991 Vol 36, No 11, pp1457 – 1464; T Szabo *et al*: "Effects of non-linearity on the

15 estimation of in-situ values of acoustic output parameters", *J Ultrasound Med* 18:33 – 41, 1999; M Hamilton *et al*: "Nonlinear acoustics", Academic Press) that the non-linearity of a medium is characterized by the coefficient of non-linearity  $\beta$ . Typical values for  $\beta$  are 3.6 for water, 4 for blood and 6.5 for fatty tissue. In addition to being nonlinear, all the media have acoustical

20 loss due to absorption. The acoustical loss is described by the power law:  $\alpha = \alpha_0 f^b$  where  $\alpha_0$  is constant and  $b$  ranges from 1 to 2 depending on the medium. For water, the rate of absorption of an ultrasound wave propagating through it is quadratically related to the frequency ( $b=2$ ). However, the rate of energy loss due to absorption is considered small and

25 most of the time the dissipation-less theory is applicable over short ranges. However, biological media have large rates of energy loss and the frequency dependence has an exponential value of 1 to 1.5. By considering both attenuation due to absorption loss and non-linearity, the exchange of energy between the two processes is complicated, because attenuation diminishes

the amplitude of the generated harmonic components with propagation distance while non-linearity builds up these harmonics. So, harmonic distortion generally tends to enrich the higher harmonic components at the expense of the lower ones (energy transfer), while absorption damps out the higher components more rapidly than the lower ones. It is therefore difficult to reach a balance in which a given component loses as much energy by absorption as it gains from nonlinear distortion. Moreover, since the conditions for stability depend on the amplitude of the wave, which slowly decreases with propagation distance, the wave can never be completely stable, only relatively so. The balance between the nonlinear process is given by the Gol'dberg number  $\Gamma$  (Szabo *et al*), which represents a measure of which likely attenuation or harmonic distortion takes over. When  $\Gamma=1$ , nonlinear effects become comparable to attenuation effects. If  $\Gamma$  is higher than 1, nonlinear processes dominate and when the Gol'dberg number is below 1, attenuation effects take over. As indication, for acoustic pressures of 500 kPa and 1MPa, at a transmit frequency of 3 MHz, the Gol'dberg number is respectively 86.5 and 43.2 for water. It is only 2.8 and 1.4 for liver-like tissue respectively at these pressures. For both settings, the parameter shows that water non-linearity is up to thirty folds higher than that of tissue.

The approach used here is based on the “non-linearity-attenuation” characteristic in differentiating between fluid media and soft tissue media. As described above, a single element transducer is placed in front of the bladder. The transducer generates a wide acoustic beam that is able to enclose the full bladder volume. Depending on the volume of urine contained in the bladder (bladder filling) and thus crossed by the acoustic beam, the amount of harmonic distortion generated in the back of the bladder will change. A radio frequency (RF) backscattered signal might be selected

from a region of interest located preferably in the backside of the bladder. The amount of energy comprised in the second harmonic or higher harmonic components of the received RF echo signal can be extracted and correlated to the amount of volume of urine that has been encompassed by the acoustic beam. Since harmonic generation is different from tissue and fluids, only the volume of urine that has been crossed by the acoustic beam would generate more harmonic energy. When the bladder is empty or below a certain volume level, no harmonic distortion occurs, whereas maximal distortion will be obtained for a full volume.

10

Figure 8 illustrates the principle of the invention. Top panel shows two situations. The bladder is either empty (Panel A left side) or filled up with urine (Panel A right side). At a certain distance beyond the bladder (around 12 cm from the transducer), a region of interest of 1.5 cm width at depth W (see figure 7) is selected. Power spectra corresponding to echo signal recorded from the regions of interest are displayed in panel B. The spectrum corresponding to the empty bladder (solid line) shows only a fundamental component. The harmonic distortion is very weak so that no harmonic frequencies are generated. However, the echo signal corresponding to the filled bladder situation (dashed line) demonstrates clear distortion where a second harmonic component with a significant energy is generated. The third harmonic component can be also present with lesser energy depending on the urine volume that has been crossed by the acoustic beam. Figure 8 demonstrates that depending on the volume contained in the bladder that the acoustic beam has intersected, the amount of generated second harmonic energy varies. When the acoustic beam crosses only tissue or when the volume of urine is very small, harmonic distortion is the lowest with no or very low harmonic energy. If the bladder is filled up or if the volume of urine is above a certain level (threshold), harmonics are generated. The

25



(  
generation of a harmonic component (2<sup>nd</sup> and/or higher harmonics) can be used for volume measurement, or simply as an indicator of filling of the bladder to a certain volume extent. The criterion can be such that if a certain amount of second harmonic (or higher harmonics) is generated in the echo  
5 signal, the device would indicate that the critical volume (or threshold) (say in adult patients around 450 ml) has been reached.

To avoid and eliminate any differences due to patient to patient variations, a normalization procedure needs to be performed a priori. Such a  
10 normalization procedure might consist of recording a first signal at very low transmit acoustic power from the same region of interest as described in the previous section. Such power would allow only linear propagation of the ultrasonic waves and avoid any harmonic generation. The echo signal would therefore have undergone only attenuation effects. In the following transmit-  
15 receive sequence, the transmit acoustic power is increased with a certain factor ( $\epsilon$ ) and a new recording is performed from the same region of interest. This measure with a much higher acoustic pressure is carried out to allow harmonic distortion to occur in the tissue. The echo signal in this case will undergo both attenuation and distortion effects. The first echo signal (linear  
20 case) will be re-scaled by the factor that corresponded to the increase in transmit power ( $\epsilon$ ), and then used as a reference signal. Consequently, each patient has his own reference hence eliminating any variations such as obesity, inhomogeneities, etc. A block diagram of a possible steps describing the second method is given in the flow chart of figure 9. The two  
25 transmitted signals might be transmitted with a very low repetition rate as indicated in figure 10. The first packet of transmit signals with low acoustic amplitude are used for calibration. The echoes received from those signals are averaged to reduce the noise level. The number of signals can be chosen such that a high signal-to-noise ratio is obtained. The second packet of

(  
signals with higher amplitudes are used to induce nonlinear propagation and harmonic distortion. The echoes received from these signals are averaged and then the harmonic energy is filtered and then compared to the calibration echo.

5

In order to estimate the volume of urine in the bladder, a look-up table can be created beforehand. Such a table, saved in the hard disk of the electronic device, will contain the correspondence between the harmonic energy and the volume of urine. Such a table can be extracted from a curve similar to  
10 the one given in figure 11. Such a curve can be obtained from a “learning” patient set of measurements. Look-up tables may eventually be produced for specific patient groups for age; gender and/or weight as an input parameter.

The described **second method** differs greatly from known other apparatus:

15

12)The device is composed of a single element defocused ultrasound transducer with a conical beam shape;

13)The single acoustic beam entirely encompasses the volumetric area of a possibly filled bladder.

20

14)The method is based on measurement of non-linear properties and attenuation behavior of propagating ultrasound waves as influenced by a urine filled bladder.

15)The method incorporates a technique to eliminate patient variation due to fat or skin properties.

25

16)The method for automatic volume computation does not assume any geometrical model for the bladder shape;

17)It is valid for any bladder shape since the received signal “integrates” all volume effects in the ultrasound beam.

(

18) All known other methods use bladder wall echoes as a basis to calculate volume.

19) The device works instantaneously.

- 5 Other embodiments are intentionally within the scope of the accompanying claims.

## CLAIMS

1. An apparatus for automatically measuring the volume of urine of a human bladder using non-invasive ultrasound echo technique in approximately the sagittal cross sectional plane, comprising: means for  
5 transmitting and receiving a plurality of ultrasound signals in a selected order, with a limited number of ultrasound transducers where the transducers with narrow sound beams are mounted under specific angles in a transducer assembly to allow estimation of volume over a wide range of volumes; means to derive a specific measurement configuration  
10 corresponding to the bladder filling degree from the number of ultrasound beams that intercept the fluid filled bladder and thus allow selection of an appropriate empirically established correction factor K for each specific measurement configuration, for optimal calculation of the volume; means to activate the transducers in a particular known  
15 sequence; means to detect the bladder wall echoes; means to use echo travel time and beam information for automatic derivation of cavity height H and depth D as part of the volume calculation depending on the specific beams that intercept the cavity; means for calculating the bladder volume according to the measurement configuration specific formula  
20  $H \times D \times K$ ; means for instantaneous display of the volume to allow to optimize transducer positioning.
2. An apparatus of claim 1, where the number of transducers can be further minimized to yield only gross ranges of bladder filling or indicate a  
25 clinically important bladder filling level.
3. An apparatus of claim 1, with means to input important patient information, for example gender, weight and age, to select the optimum

(  
set of empirically established set of correction factors for the volume calculation.

4. An apparatus of claim 1, provided with internal memory and means to  
5 input and store validated volume measurements to optimize correction factors in a "self learning process".

5. An apparatus of claim 1, provided with means to freeze the read-out with a hold/start button.

10

6. An apparatus of claim 1, where the transducers are positioned in an assembly so that the echo reflecting area's of the walls of the bladder are approximately located in a single cross-sectional sagittal plane.

15 7. An apparatus of claim 1, where the shape of the used ultrasound transducers is approximately a disk.

8. An apparatus of claim 1, where the electric power can be provided by a battery.

20

9. An apparatus of claim 1, where the correct caudal-cranial positioning of the transducer assembly over the bladder is indicated on the transducer assembly.

25 10. A method for non-invasive automatic measurement and calculation of bladder volume comprising the following steps: transmitting and receiving a plurality of ultrasonic waves; automatically detecting the number of beams that intercept the bladder; automatically processing the

(  
bladder height and depth; determining the volume of urine by applying a specific correction factor; displaying the volume computed.

- 5 11. An apparatus of claim 1, where the transducer is connected with a cable to a housing containing the input-electronic components-calculation-display-power supply unit.
12. An apparatus of claim 1, where housing and transducer assembly presented in claim 11 are integrated.
- 10 13. An apparatus of claim 1, where the transducer assembly at the transducer-skin interface is provided with ultrasound coupling rubber material for optimal acoustic coupling and patient convenience.
- 15 14. An apparatus for automatically and non-invasively measuring the urine volume of the human bladder using an ultrasound echo technique comprising: means for transmitting and receiving ultrasound signals with a single transducer with a fundamental transmission frequency in a wide sound beam that intercepts the bladder area; means for selecting the received ultrasound signal from a given depth range; means for measuring the presence of higher harmonics of the transmitted fundamental frequency in the received echo signal from that given depth range; means to convert this signal into a measurement of the urine filling of the bladder.
- 20 25 15. An apparatus of claim 14, where the ultrasound transducer is a curved single active piezo-electric element, shaped to form a sector of a sphere or cone like sound beam to approximately encompass a filled bladder.

- (
16. An apparatus of claim 14, where the ultrasound transducer is combined with a lens material so that the combination of transducer and lens create a wide sound beam to approximately encompass the filled bladder.
  - 5 17. An apparatus of claim 14, where the transducer is capable to transmit at a fundamental ultrasound frequency and is capable to receive the fundamental and higher harmonic signals of the transmitted frequency.
  18. An apparatus of claim 14, where in transmission at the fundamental  
10 frequency a multiplicity of different pulses can be transmitted to in combination enhance the non-linear presence in the received echoes.
  19. An apparatus of claim 14, where the received signal can be selected from approximately a depth or distance which is beyond the posterior wall of  
15 an average human filled bladder.
  20. An apparatus of claim 14, where the scattered power of higher harmonics in the received signal can be derived and where these signals can be compared with the backscattered power in the fundamental frequency  
20 and used in an algorithm to calculate the urine volume.
  21. An apparatus of claim 14, where combined pulse sequences at low transmit power and high transmit power are used to enhance bladder filling measurement and eliminate patient variation due to for instance  
25 obesity.
  22. An apparatus of claim 14, where in the volume algorithm also echo data from a depth close to the position of the average anterior bladder wall is incorporated to limit patient variation proximal to the transducer.

23. An apparatus of claim 14, where at various depth by changing the transmitted power in subsequent pulse transmission, linear and non-linear echo signals can be compared to eliminate patient variation.

5

24. An apparatus of claim 14, where the volume calculation is based on a prior established look up table.

25. An apparatus of claim 14, where the transducer, the electronics, the rechargeable battery power and the volume display are combined in a single housing.

10

26. An apparatus of claim 14, where the display is limited to indicate filling above a certain critical level given the medical application.

15

27. An apparatus of claim 14, where the display is limited to indicate filling below a certain critical level given the medical application.

20

28. An apparatus of claim 14, where the transducer is connected with a flexible cable to the apparatus comprising the electronics and electric power.

25

29. An apparatus of claim 14, where the transducer is essentially a combination of an acoustic active surface for optimal transmission and reception at the fundamental frequency and another acoustic active surface for optimal reception of the higher harmonic echo signals.

30. An apparatus of claim 14, where a transducer of the apparatus described in claim 1 with narrow beam direction in dorsal direction is used to detect



the anterior and posterior bladder wall; this information is used for appropriate selection of the echo depth for proper recording of the higher harmonic echo signal.

- 5 31. Apparatus substantially as described herein with reference to the accompanying drawings.



INVESTOR IN PEOPLE

**Application No:** GB 0218547.8  
**Claims searched:** 14-30

**Examiner:** Jacob Collins  
**Date of search:** 17 February 2003

## Patents Act 1977 : Search Report under Section 17

### Documents considered to be relevant:

Category	Relevant to claims	Identity of document and passage or figure of particular relevance
A		EP 0271214 A2 (NASA)
A		US 6213949 B1 (GANGULY ET AL)
A		WO 00/71030 A1 (DIAGNOSTIC ULTRASOUND CORPORATION)

### Categories:

X Document indicating lack of novelty or inventive step	A Document indicating technological background and/or state of the art.
Y Document indicating lack of inventive step if combined with one or more other documents of same category.	P Document published on or after the declared priority date but before the filing date of this invention.
& Member of the same patent family	E Patent document published on or after, but with priority date earlier than, the filing date of this application.

### Field of Search:

Search of GB, EP, WO & US patent documents classified in the following areas of the UKC<sup>v</sup>:

G1G

Worldwide search of patent documents classified in the following areas of the IPC<sup>7</sup>:

A61B

The following online and other databases have been used in the preparation of this search report :

WPI, EPODOC, JAPIO



INVESTOR IN PEOPLE

**Application No:** GB 0218547.8  
**Claims searched:** 1-13

**Examiner:** Jacob Collins  
**Date of search:** 20 December 2002

## Patents Act 1977 : Search Report under Section 17

### Documents considered to be relevant:

Category	Relevant to claims	Identity of document and passage or figure of particular relevance	
X	1,2,4,10,13	US 6359190 B1	(TER-OVANEXYAN ET AL) column 3 lines 38-67 and column 10 lines 14 -40 in particular
A	-	US 4926871	(GANGULY ET AL)
A	-	US 5964710	(GANGULY ET AL)

### Categories:

X	Document indicating lack of novelty or inventive step	A	Document indicating technological background and/or state of the art.
Y	Document indicating lack of inventive step if combined with one or more other documents of same category.	P	Document published on or after the declared priority date but before the filing date of this invention.
&	Member of the same patent family	E	Patent document published on or after, but with priority date earlier than, the filing date of this application.

### Field of Search:

Search of GB, EP, WO & US patent documents classified in the following areas of the UKC<sup>1</sup>:

G1G

Worldwide search of patent documents classified in the following areas of the IPC<sup>7</sup>:

A61B

The following online and other databases have been used in the preparation of this search report:

WPI, EPODOC, JAPIO